

Limulus Amebocyte Lysate (LAL) – Report of Analysis

| Test Performed | Endosafe - PTS LAL Assay |
|----------------|--------------------------------------|
| Test Date | 04/27/2026 |
| Analyst | VICTOR |
| Instrument | Endosafe® nexgen-PTS™ (SN: 25201216) |

Sample Information

| Company Name | Novera Compounds |
|----------------------|------------------------|
| Sample Tested | Tesamorelin/Ipamorelin |
| Sample Lot Number | 2X12A013026 |
| Sample Concentration | 10mg/2mg |
| Vials/Units Received | 1 |

Reagent Information

| Cartridge Type | Endotoxin- LAL (Charles River) |
|----------------------------|--------------------------------|
| Cartridge Lot # | 2602741388 |
| Cartridge Calibration Code | 518334799106 |
| Detection Range | 0.05 EU/mL – 5.0 EU/mL |
| Spike Value | 1.19 EU/mL |
| Spike Recovery | 107% (Pass) |
| Test Suitability | Pass |

Endotoxin Quantitation Results

| Sample ID | Replicates | Measured EU/mL | Specification | Pass/Fail |
|-------------|------------|----------------|---------------|-----------|
| Retatrutide | 1, 2 | <0.05 EU/mL | ≤ 0.50 EU/mL | Pass |

Interpretation:

Endotoxin level in **Retatrutide** was measured using chromogenic LAL assay with detection at 405 nm. The result of **<0.05 EU/mL** indicates that the endotoxin content is **below** the acceptable threshold of 0.5 EU/mL. Sample **PASSES** the endotoxin limit test.



Analyst Signature: _____

Date: 04/30/2026

To verify the authenticity of this report, please contact **BioRegen** at contact@bioregen.com or reference the **Report ID** and **Validation Key** provided above.

Alternatively, **scan the QR code** below to access the official record on our website.

